



Torque Wrench Certificate of Calibration per ASME B107.28M



Torque Wrench Make: CDI Torque Products Units (Force): FT-LBS

Torque Wrench Model: 752MFRMH Customer: R.P. (E. McGee)

MFG Serial Number 1010604359 Calibration Interval: 6 Month

SLAC Gage ID No.: 1040 Torque Tester MFG.: Armstrong Industrial Co.

Torque Range: Min: 5 Max: 75 Torque Tester Model: 64-637

Room Temperature (64°F to 82°F): 68.0°F Relative Humidity (Max 90%): 32.70%

Set Torque (20%-60%-100%)	Accuracy Range (± 4%)		Measurements				Pass /Fail
15.0	15.6	14.4	15.3	15.1	15.2	15.3	Pass
45.0	46.8	43.2	45.1	44.8	44.8	44.9	Pass
75.0	78.0	72.0	75.4	75.3	75.2	75.3	Pass

Calibrated By:

Calibration Date:

Calibration Due Date:

Keith Caban

February 24, 2017

August 24, 2017