

SLAC Traveler for LCLS undulator re-tuning.

Undulator S/N # 09 Dataset0002

This traveler is intended to cover magnetic measurements and mechanical fiducialization of the undulator segments at MMF.

Preparation:

Move the undulator inside the temperature controlled room and keep it in the crate for 7 days to get the room temperature.

Placing undulator on the measurement bench:

Earthquake T-bolts checked	XRD	✓
Thermistor blocks attached	XRD	✓
X-trajectory shims modified	XRD	✓
Y-trajectory shims replaced		✓
Check interference with probes	XRD	✓
Check end plates	XRD	✓
Technician (initials):	XRD	
Date (mm-dd-yyyy):	04/20/2009	

Tuning.

Follow the fine tuning test plan to align the undulator to the bench, measure x , y field and calculate the trajectories, phases and field integrals, and fiducialize the undulator (LCLS-TN-06-17, LCLS-TN-07-2). X and Y to be set to 10 μ m; roll to 0.1mrad; pitch to 0.005mrad, and yaw to 0.010mrad.

Mechanical alignment done	✓
Magnetic alignment done	✓
Probe roll angle checked	✓

The following information is to be noted by an engineer upon finishing alignment to the granite:

Engineer (initials):	Y
Date (mm-dd-yyyy):	04/20/09
Average X (m):	0.029133
Average Y (m):	0.000419
Final Roll (rad):	$-2 \cdot 10^{-6}$
Final Pitch (rad):	$+0.7 \cdot 10^{-6}$
Final Yaw (rad):	$+1.2 \cdot 10^{-6}$

$$t = 19.98^{\circ}\text{C}$$

The following information is to be noted by an engineer upon final tuning:

Engineer (Initials):	Y
Date (mm/dd/yyyy)	04/22/09
Reference magnet measurement (T):	-0.369154
First integral Y at center (Tm):	$-9 \cdot 10^{-6}$
Second integral Y at center (Tm ²):	$-18 \cdot 10^{-6}$
First integral X at center (Tm):	$-1 \cdot 10^{-6}$
Second integral X at center (Tm ²):	$-4 \cdot 10^{-6}$
Magnetic axis position X (m):	0.029117
Magnetic axis position Y (m):	0.000434
Final Phase error rms (°):	3.3
Final max X trajectory error (μm):	-0.6
Final max Y trajectory error (μm):	-0.5
Measured K-value:	3.497221
X-field correction (T):	-0.21
Y-field correction (T):	-0.05
Final Phase error entrance (°):	-0.9
Final Phase error cell (°):	-1.0
Final Phase error exit (°):	+2.5
Slot number	7

URL of on-line Final tuning data:

www-group.slac.stanford.edu/met/MagMeas/MAGDATA/LCLS/Undulator/L143-112000-09/DATASET002/Fine%20Tuning/

Fiducialization on Kugler bench:

Upon completion of the tuning the undulator should be fiducialized; optical measurements to be done the same day as magnetic measurements.

Attach pointed magnets to the both ends of the undulator. Measure position of the PM centers w.r.t. the undulator center line.

The following information is to be noted by an engineer:

Engineer (Initials):	X
Upstream PM X offset (m):	0.029340
Upstream PM Y offset (m):	0.000342
Downstream PM X offset (m):	0.029273
Downstream PM Y offset (m):	0.000266
Reference upstream PM X offset (m):	0.076841
Reference upstream PM Y offset (m):	0.000326
Reference downstream PM X offset (m):	0.076802
Reference downstream PM Y offset (m):	0.000317
Keyence probe (m):	+0.011147
Keyence block (m):	-0.011367

Have an alignment crew to measure offsets between PM tooling balls, the undulator tooling balls, reference PM tooling balls to an alignment reference line, Hall probe in X and gage block in Y. Attach alignment data sheet to the traveler.


Surveyor (initials):	KG
Date (mm/dd/yyyy):	04/22/09

Move the undulator to CMM to finalize the fiducialization.

Attach CMM data sheet to the traveler.

URL of on-line fiducialization data:

www-group.slac.stanford.edu/met/MagMeas/MAGDATA/LCLS/Undulator/L143-112000-09/DATASET002

Undulator tuning completed (signed):	
Date (mm/dd/yy)	04/27/09