# SLAC Traveler for LCLS undulator re-tuning. Undulator S/N # 20 Dataset0005

This traveler is intended to cover magnetic measurements and mechanical fiducialization of the undulator segments at MMF.

## Preparation:

Move the undulator inside the temperature controlled room and keep it in the crate for 7 days to get the room temperature.

#### Placing undulator on the measurement bench:

Earthquake T-bolts checked		RC
Thermistor blocks attached		& c
X-trajectory shims modified		1./4 ( \
Y-trajectory shims replaced		M/#-{
Check interference with probes		
Check end plates	š	PC
Technician (initials):		20
Date (mm-dd-yyyy):		•

6/4/00

# Tuning.

Follow the fine tuning test plan to align the undulator to the bench, measure x , y field and calculate the trajectories, phases and field integrals, and fiducialize the undulator (LCLS-TN-06-17, LCLS-TN-07-2). X and Y to be set to  $10\mu m;$  roll to 0.1mrad; pitch to 0.005mrad, and yaw to 0.010mrad.

Mechanical alignment done	V
Magnetic alignment done	V
Probe roll angle checked	マ

The following information is to be noted by an engineer upon finishing alignment to the granite:

Engineer (initials):	X
Date (mm-dd-yyyy):	06/04/09
Average X (m):	0.02838.00
Average Y (m):	0.000516
Final Roll (rad):	+1.10-6
Final Pitch (rad):	- Q. 6·10 <sup>-6</sup>
Final Yaw (rad):	+0.4.10-6

t=20.00 °C

The following information is to be noted by an engineer upon final tuning:

Engineer (Initials):	X
Date (mm/dd/yyyy)	06/05/09
Reference magnet measurement (T):	06/05/09 -0.369/5/ -7.10 <sup>-6</sup>
First integral Y at center (Tm):	-7·10 <sup>-6</sup>
Second integral Y at center (Tm <sup>2</sup> ):	0
First integral X at center (Tm):	0
Second integral X at center (Tm <sup>2</sup> ):	$-19.10^{-6}$
Magnetic axis position X (m):	0.029166
Magnetic axis position Y (m):	0.000524
Final Phase error rms (°):	3.5
Final max X trajectory error (μm):	+0.6
Final max Y trajectory error (µm):	-0.3
Measured K-value:	3.485339
X-field correction (G):	-0.14
Y-field correction (G):	-0.06
Final Phase error entrance (°):	+1.3
Final Phase error cell (°):	+2.5
Final Phase error exit (°):	-1.9
Slot number	33

URL of on-line Final tuning data:

www-group.slac.stanford.edu/met/MagMeas/MAGDATA/LCLS/Undulator/L143-112000-20/DATASET005/Fine%20Tuning/

### Fiducialization on Kugler bench:

Upon completion of the tuning the undulator should be fiducialized; optical measurements to be done the same day as magnetic measurements.

Attach pointed magnets to the both ends of the undulator. Measure position of the PM centers w.r.t. the undulator center line.

The following information is to be noted by an engineer:

Engineer (Initials):	X
Upstream PM X offset (m):	0.029226
Upstream PM Y offset (m):	0.000452
Downstream PM X offset (m):	0.029184
Downstream PM Y offset (m):	0.000366
Reference upstream PM X offset (m):	0.076927
Reference upstream PM Y offset (m):	0.000426
Reference downstream PM X offset (m):	0.076894
Reference downstream PM Y offset (m):	0.000428
Keyence probe (m):	+0.0/1/35
Keyence block (m):	-0.011364

Have an alignment crew to measure offsets between PM tooling balls, the undulator tooling balls, reference PM tooling balls to an alignment reference line, Hall probe in X and gage block in Y. Attach alignment data sheet to the traveler.

Surveyor (initials):	LG
Date (mm/dd/yyyy):	06/05/09

Move the undulator to CMM to finalize the fiducialization.

Attach CMM data sheet to the traveler.

URL of on-line fiducialization data:

 $\underline{www-group.slac.stanford.edu/met/MagMeas/MAGDATA/LCLS/Undulator}$ 

/L143-112000-20/DATASET005

·	
Undulator tuning completed (signed):	whelearen s
Date (mm/dd/yy)	06/08/09