

```

=====
Routine Name                               Run #       Date & Time
=====
PF_235_910_58.RTN                          1 Mon Sep 13 10:14:14 2004
=====

```

```

Stanford Linear Accelerator Center
WR62 WG Flange Adapter
PF-235-910-58
QC'd by: T. Nakashima
SN: 001
=====

```

Feature	Unit	Nominal	Actual	Tolerances	Deviation	Exceeds
Step 4						
X Location	in	+00.81800	+00.81579	+0.00000 -0.00500	-00.00221	--
Step 5						
Y Location	in	+00.50800	+00.50479	+0.00000 -0.00500	-00.00321	---
Step 12						
Width	in	+00.33500	+00.33927	+0.00200 -0.00200	+00.00427	+00.00227
Y Location	in	+00.00000	+00.00447	+0.00150 -0.00150	+00.00447	+00.00297
Step 15						
Width	in	+00.64600	+00.64672	+0.00200 +0.00000	+00.00072	++
X Location	in	+00.00000	-00.00023	+0.00150 -0.00150	-00.00023	-
Step 20						
Width	in	+00.68800	+00.68825	+0.00200 -0.00200	+00.00025	+
X Location	in	+00.00000	-00.00005	+0.00150 -0.00150	-00.00005	-
Step 23						
Width	in	+00.37700	+00.37670	+0.00500 -0.00500	-00.00030	-
Y Location	in	+00.00012	+00.00012		-00.00000	
Step 26						
Z Location	in	-00.20000	-00.20168	+0.00500 -0.00500	+00.00168	++
Step 28 - Z = 4X .075						
Z Location	in	+00.07500	+00.07712	+0.00500 -0.00500	+00.00212	++
Step 29 - Flatness						
Flatness	in		+00.00012	+0.00200	+00.00012	+

PARALLEL 001 A

0002