

Quindos Measuring Report



3D-Application Center

Description	STARRETT MIC STAND	Customer	0
Drawing Number		Serial Number	STAR-MIC-STD-006-01
Remarks		Article Number	
Supplier	STARRETT	Delivery Date	
Delivery Note		Delivery Volume	
Lot Number		Lot Size	
Test Schedule		Sample Size	
Production Machine		Production Tool	
Production Date		Production Time	
Order	0	Department	MET. QUALITY INSPEC
Inspector	kcaban	Inspection Date	29-APR-2009, 08:56:34
Measuring Device	PMM 12 10 6 #496	Measuring Program	Quindos7 - V 7.1.9119-B
User Name	kcaban	WKP Name	CAL_STD

Text	Eval.	Actual	Nominal	Up.To1.	Low.To1.	Act-Nom	Graphic
PT							POI
	Z	-6.00003	6.00000	0.00010	-0.00010	0.00003	
PT(1)							POI
	Z	-6.00003	6.00000	0.00010	-0.00010	0.00003	
PT(2)							POI
	Z	-6.00003	6.00000	0.00010	-0.00010	0.00003	
PT(3)							POI
	Z	-6.00003	6.00000	0.00010	-0.00010	0.00003	
PT(4)							POI
	Z	-6.00003	6.00000	0.00010	-0.00010	0.00003	
PT(5)							POI
	Z	-6.00003	6.00000	0.00010	-0.00010	0.00003	
PT(6)							POI
	Z	-6.00003	6.00000	0.00010	-0.00010	0.00003	
PT(7)							POI
	Z	-6.00003	6.00000	0.00010	-0.00010	0.00003	
PT(8)							POI
	Z	-6.00003	6.00000	0.00010	-0.00010	0.00003	
PT(9)							POI
	Z	-6.00003	6.00000	0.00010	-0.00010	0.00003	
PT(10)							POI
	Z	-6.00003	6.00000	0.00010	-0.00010	0.00003	
STATS							
	MEAN	-6.00003	6.00000	0.00010	-0.00010	0.00003	
	MIN	-6.00003	6.00000	0.00010	-0.00010	0.00003	
	MAX	-6.00003	6.00000	0.00010	-0.00010	0.00003	
	RANGE	0.00000	0.00000			0.00000	
	SIGMA	0.00000	0.00000			0.00000	

Text	Eval.	Actual	Nominal	Up.To1.	Low.To1.	Act-Nom	Graphic
Date: 29-Apr-09, 08:56:55							Page 1 of 1
						CAL_STD	