

Y Value Scales Centered on Mean Values  
Step Between Measured Pole Tips = 1

Regression Line Through Points = -----  
Dimensions in mm




METROLOGY

# LCLS II-HE - SXR Undulator

Nominal Gap = 10      Nominal Taper = 0.000  
Gap Reading = 9.9997      US Encoder = 9.9994      DS Encoder = 10.0000

15-JAN-2025  
S/N = 029  
D/S = 0001  
Run = 4

# Gap Between Opposing Pole Tips



Step Between Measured Pole Tips = 1

Dimensions in mm




METROLOGY

## LCLS II-HE - SXR Undulator

Nominal Gap = 10  
Nominal Taper = 0.000  
Gap Reading = 9.9997  
US Encoder = 9.9994  
DS Encoder = 10.0000

15-JAN-2025  
S/N = 029  
D/S = 0001  
Run = 4

### Symmetry Point Between Pole Tips



Mean Symmetry Value = -0.1964

Step Between Measured Pole Tips = 1

Dimensions in mm



METROLOGY

## LCLS II-HE - SXR Undulator

Nominal Gap = 10      Nominal Taper = 0.000  
Gap Reading = 9.9997      US Encoder = 9.9994      DS Encoder = 10.0000

15-JAN-2025  
S/N = 029  
D/S = 0001  
Run = 4

## Top and Bottom Jaw Regression Line Intersect Points

Jaw	First Pole <i>(Pole -1)</i>	US Actuator <i>(Pole 28)</i>	DS Actuator <i>(Pole 90)</i>	Last Pole <i>(Pole 118)</i>
<b>Top</b>	4.7969	4.7960	4.7940	4.7931
<b>Bottom</b>	-5.1842	-5.1858	-5.1896	-5.1912
<b>Gap</b>	9.9811			9.9843
<b>Taper</b>				0.0032

### Summary of Mean Values

Top Jaw Poles	Btm. Jaw Poles	Gap Values	Sym. Pt. Values
4.7950	-5.1877	9.9827	-0.1964

### Additional Calculated Values

<b>Bottom Pole #1 Z Value</b>	<b>978.618</b>
<b>Top Jaw Pitch (mrad)</b>	<b>-0.001</b>
<b>Bottom Jaw Pitch(mrad)</b>	<b>-0.002</b>
<b>Minimum Effective Gap</b>	<b>9.962</b>
<b>Reference Block Gap</b>	<b>6.794</b>

Dimensions in mm



METROLOGY

## LCLS II-HE - SXR Undulator

Nominal Gap = 10      Nominal Taper = 0.000  
 Gap Reading = 9.9997      US Encoder = 9.9994      DS Encoder = 10.0000

15-JAN-2025  
 S/N = 029  
 D/S = 0001  
 Run = 4