


Minimum Effective Gap = 9.966

Y Value Scales Centered on Mean Values
 Step Between Measured Pole Tips = 1

Regression Line Through Points = -----
 Dimensions in mm

	<h2 style="margin: 0;">LCLS II - SXR Undulator</h2> <p style="margin: 0;">Nominal Gap = 10</p> <p style="margin: 0;">Gap Reading = 10.0000 US Encoder = 10.0000 DS Encoder = 10.0000</p>	<p style="margin: 0;">07-MAR-2018</p> <p style="margin: 0;">S/N = 007</p> <p style="margin: 0;">D/S = 0001</p> <p style="margin: 0;">Run = 5</p>
---	--	--

Gap Between Opposing Pole Tips



Step Between Measured Pole Tips = 1

Dimensions in mm



LCLS II - SXR Undulator

Nominal Gap = 10

Gap Reading = 10.0000 US Encoder = 10.0000 DS Encoder = 10.0000


07-MAR-2018

S/N = 007

D/S = 0001

Run = 5

Symmetry Point Between Pole Tips



Mean Symmetry Value = 0.2373

Step Between Measured Pole Tips = 1

Dimensions in mm



LCLS II - SXR Undulator

Nominal Gap = 10

Gap Reading = 10.0000 US Encoder = 10.0000 DS Encoder = 10.0000

07-MAR-2018
 S/N = 007
 D/S = 0001
 Run = 5

Top and Bottom Jaw Regression Line Intersect Points

Jaw	First Pole (Pole 1)	US Actuator (Pole 39)	DS Actuator (Pole 135)	Last Pole (Pole 174)
Top	5.2464	5.2429	5.2339	5.2303
Bottom	-4.7636	-4.7637	-4.7639	-4.7640
Gap	10.0100			9.9943
Taper				-0.0158

Summary of Mean Values

Top Jaw Poles	Btm. Jaw Poles	Gap Values	Sym. Pt. Values
5.2383	-4.7638	10.0021	0.2373

Additional Calculated Values

Bottom Pole #1 Z Value	979.176
Top Jaw Pitch (mrad)	0.005
Bottom Jaw Pitch(mrad)	0.000
Minimum Effective Gap	9.966
Reference Block Gap	6.804

Dimensions in mm



LCLS II - SXR Undulator

Nominal Gap = 10

Gap Reading = 10.0000 US Encoder = 10.0000 DS Encoder = 10.0000

07-MAR-2018
S/N = 007
D/S= 0001
Run= 5