


Minimum Effective Gap = 9.982

Y Value Scales Centered on Mean Values
Step Between Measured Pole Tips = 1

Regression Line Through Points = -----
Dimensions in mm

	LCLS II - SXR Undulator		28-MAR-2018
	Nominal Gap = 10	Nominal Taper = 0.1	S/N = 008
	Gap Reading = 10.0000	US Encoder = 10.0000	DS Encoder = 10.0000
			Run= 5



Step Between Measured Pole Tips = 1


Dimensions in mm



LCLS II - SXR Undulator

Nominal Gap = 10 Nominal Taper = 0.1
 Gap Reading = 10.0000 US Encoder = 10.0000 DS Encoder = 10.0000

28-MAR-2018
 S/N = 008
 D/S = 0001
 Run = 5



Mean Symmetry Value = 0.1059

Step Between Measured Pole Tips = 1

Dimensions in mm



LCLS II - SXR Undulator

Nominal Gap = 10 Nominal Taper = 0.1
 Gap Reading = 10.0000 US Encoder = 10.0000 DS Encoder = 10.0000

28-MAR-2018
 S/N = 008
 D/S = 0001
 Run = 5

Top and Bottom Jaw Regression Line Intersect Points

Jaw	First Pole (Pole 1)	US Actuator (Pole 39)	DS Actuator (Pole 135)	Last Pole (Pole 174)
Top	5.0944	5.1063	5.1362	5.1484
Bottom	-4.9185	-4.9146	-4.9048	-4.9008
Gap	10.0129			10.0492
Taper				0.0363

Summary of Mean Values

Top Jaw Poles	Btm. Jaw Poles	Gap Values	Sym. Pt. Values
5.1214	-4.9096	10.0310	0.1059

Additional Calculated Values

Bottom Pole #1 Z Value	979.390
Top Jaw Pitch (mrad)	0.016
Bottom Jaw Pitch(mrad)	0.005
Minimum Effective Gap	9.982
Reference Block Gap	6.803

Dimensions in mm



LCLS II - SXR Undulator

Nominal Gap = 10 Nominal Taper = 0.1
 Gap Reading = 10.0000 US Encoder = 10.0000 DS Encoder = 10.0000

28-MAR-2018
 S/N = 008
 D/S= 0001
 Run= 5