SLAC Magnetic Measurements

Planar Insertion Device Shim Analysis

SLAC Magnetic Measurements
Date: 08−12−2013
Time: 10:26:54

Project: LCLS
Device Type: Undulator
Data Type:
Serial Number: L143−112000−35
DATASET Number: 5
Tuning Stage: Fine Tuning
Run Number: 7
Measurement Type: Z Scans
Operator: YL
Comment: calibration Run 80, check K
Session first run: 7
Measurement Device x:
Measurement Device y: $SENTRON_4$41

Data File: s:\magdata\LCLS\Undulator\L143−112000−35\DATASET0005\Fine Tuning\Z Scans\Run 007\bscanz.dat

Temperature (C)
L143−112000−35:−−> T1 20.00 −−> T2 19.99 −−> T5 19.93 −−> T3 19.99 −−> Air 19.98 −−> T4 19.97

Axes position (m) X:0.029663 Y:0.000211
Initial Hall Probe Zero Measurements:
Vx_0i = 0.000947 V
Vy_0i = −0.003021 V



Phase Error: Ave = −0.000000 deg, RMS = 3.400463 deg