SLAC Magnetic Measurements

Planar Insertion Device Analysis

SLAC Magnetic Measurements
Date: 04−14−2014
Time: 08:31:05

Project: LCLS
Device Type: Undulator
Data Type:
Serial Number: L143−112000−05
DATASET Number: 3
Tuning Stage: Fine Tuning
Run Number: 4
Measurement Type: Z Scans
Operator: YL
Comment: check roll
Session first run: 2
Measurement Device x:
Measurement Device y: $SENTRON_4$41

Data File: s:\magdata\LCLS\Undulator\L143−112000−05\DATASET0003\Fine Tuning\Z Scans\Run 004\bscanz.dat

Temperature (C)
L143−112000−05:−−> T1 20.03 −−> T2 19.99 −−> T5 19.94 −−> T3 19.99 −−> Air 19.90 −−> T4 19.96

Axes position (m) X:0.030011 Y:0.000200
Initial Hall Probe Zero Measurements:
Vx_0i = 0.001209 V
Vy_0i = −0.003110 V


Bx vs Z
x 10^-3
13
46
80
113
146
180
213
Pole
Bx (T)
Z Position (m)
4
3
2
1
0
−1
−2
−3
−2.5
−2
−1.5
−1
−0.5
0
0.5
1
1.5
2
2.5


By vs Z
13
46
80
113
146
180
213
Pole
1.5
1
0.5
0
−0.5
−1
−1.5
By (T)
−2.5
−2
−1.5
−1
−0.5
0
0.5
1
1.5
2
2.5
Z Position (m)


Delta By Peaks vs Pole Number
By (G)
Pole Number
100
0
−100
−200
−300
−400
−500
0
20
40
60
80
100
120
140
160
180
200


X Trajectory
x 10^-6
13
46
80
113
146
180
213
Pole
2
1.5
1
0.5
0
−0.5
−1
−1.5
−2
X (m)
−2.5
−2
−1.5
−1
−0.5
0
0.5
1
1.5
2
2.5
Z position (m)


Y Trajectory
x 10^-6
0
−0.5
−1
−1.5
−2
−2.5
−3
−3.5
−4
Y (m)
13
46
80
113
146
180
213
Pole
−2.5
−2
−1.5
−1
−0.5
0
0.5
1
1.5
2
2.5
Z position (m)



Phase Error: Ave = 0.000000 deg, RMS = 4.183489 deg


Phase Error Relative To Nominal Radiation Wave At Field Peaks, No Ends
13
46
80
113
146
180
213
Pole
30
25
20
15
10
5
0
−5
−10
Phase Error (deg)
−2
−1.5
−1
−0.5
0
0.5
1
1.5
2
Z position (m)