SLAC Magnetic Measurements

Planar Insertion Device Analysis

SLAC Magnetic Measurements
Date: 02−06−2014
Time: 09:25:41

Project: LCLS
Device Type: Undulator
Data Type:
Serial Number: L143−112000−25
DATASET Number: 3
Tuning Stage: Fine Tuning
Run Number: 11
Measurement Type: Z Scans
Operator: YL
Comment: find Y at min K; before final
Session first run: 11
Measurement Device x:
Measurement Device y: $SENTRON_4$41

Data File: s:\magdata\LCLS\Undulator\L143−112000−25\DATASET0003\Fine Tuning\Z Scans\Run 011\bscanz.dat

Temperature (C)
L143−112000−25:−−> T1 19.92 −−> T2 19.90 −−> T5 19.83 −−> T3 19.91 −−> Air 19.82 −−> T4 19.87

Axes position (m) X:0.029996 Y:0.000137
Initial Hall Probe Zero Measurements:
Vx_0i = 0.001202 V
Vy_0i = −0.003287 V


Bx vs Z
x 10^-3
5
0
−5
−10
13
46
80
113
146
180
213
Pole
Bx (T)
−2.5
−2
−1.5
−1
−0.5
0
0.5
1
1.5
2
2.5
Z Position (m)


By vs Z
13
46
80
113
146
180
213
Pole
1.5
1
0.5
0
−0.5
−1
−1.5
By (T)
−2.5
−2
−1.5
−1
−0.5
0
0.5
1
1.5
2
2.5
Z Position (m)


Delta By Peaks vs Pole Number
By (G)
Pole Number
100
0
−100
−200
−300
−400
−500
−600
0
20
40
60
80
100
120
140
160
180
200


X Trajectory
x 10^-6
13
46
80
113
146
180
213
Pole
2
1.5
1
0.5
0
−0.5
−1
−1.5
−2
X (m)
−2.5
−2
−1.5
−1
−0.5
0
0.5
1
1.5
2
2.5
Z position (m)


Y Trajectory
x 10^-6
13
46
80
113
146
180
213
Pole
2
1.5
1
0.5
0
−0.5
−1
−1.5
−2
Y (m)
−2.5
−2
−1.5
−1
−0.5
0
0.5
1
1.5
2
2.5
Z position (m)



Phase Error: Ave = −0.000000 deg, RMS = 2.810325 deg


Phase Error Relative To Nominal Radiation Wave At Field Peaks, No Ends
13
46
80
113
146
180
213
Pole
8
6
4
2
0
−2
−4
−6
−8
−10
−12
Phase Error (deg)
−2
−1.5
−1
−0.5
0
0.5
1
1.5
2
Z position (m)