

**Jobe, R. Keith**

---

**From:** Bruce McCaul [BruceM@oxigraf.com]  
**Sent:** Friday, May 23, 2008 10:48 AM  
**To:** Jobe, R. Keith  
**Cc:** alane@oxigraf.com; 'Tom Winsemius'; jasonm@oxigraf.com  
**Subject:** Oxigraf  
**Attachments:** 08-0427-B0, Manual, O2iM.pdf; Safety Monitor Brochure v1-web.pdf; 03-0043 A11 Price List O2iM.pdf; Assembly, Indicator, SemiStat.pdf; indicator, box, remote, 4-channel.jpg; Wiring Instructions, Remote Indicator, 4-Color.pdf

Keith Jobe  
SLAC  
650 926 2084  
650 215 8084 cell

Keith,

Thanks for your call today! We understand that you are Chairman of the SLAC Safety Committee and will be taking over some of the roles that Ron Badger filled in the past.

The remote box indicator is shown for 4-port “channel relay” mode installations: see Manual 3.1.6. This could be made with blue LEDs. We can also supply this style remote box with the 4-color “Normal” mode lights.

Attached is further information about Oxigraf oxygen safety monitors. You can learn more from our Service Manager, Jason McCartney, and from Tom Winsemius, our Senior Mechanical Engineer.

We look forward to receiving an order for an additional monitor—Alan Edelman is the Office Manager who handles our order processing.

Regards,  
Bruce

Bruce W. McCaul, Ph.D.  
bmccaul@oxigraf.com  
www.oxigraf.com  
650 237 0155 x221